



PHARMA DEVILS

QUALITY ASSURANCE DEPARTMENT

QUALITY RISK ASSESSEMENT AND MITIGATION PLAN

QRA No.:

Name of Facility/Equipment/Utility/System/Activity/Procedure
Risk in reducing Swab & rinse testing

Date of Quality Risk Assessment:

Sr. No.	Item/ Function	Potential Failure Mode (Failure Mode)	Potential Effect of Failure(Effect)	Potential Cause/ Mechanism of Failure	Current Control	Reference	S	O	D	Risk Priority Number (S*O*D)	Recommend- ended Actions (if any)	Post Risk			
												S	O	D	RP N S*O
1.	Cleaning Verification	Rinse Sample not send to Quality control Dept.after each product to product change over.	May leads to contamination	<ul style="list-style-type: none"> ➤ Product Failure ➤ Product recall ➤ Product quality 	<ul style="list-style-type: none"> ➤ Cleaning process has been validated for practically insoluble API (Nandrolon Decanoate Injection in Ampoule line, Ondansetron injection in Liquid Vial line Hydrocortisone Sodium Succinate Injection in Dry powder injection) ➤ Cleaning validation already done for Ampoule ,Liquid Vial & Dry Powder injection. ➤ Cleaning method is robust (Previous trend of swab & rinse). ➤ Visual inspection is the part of line clearance. ➤ Testing done online at shop floor and instrument is dedicated ➤ pH & Conductivity test shall be done by IPQA . 	<ul style="list-style-type: none"> ➤ SOP No.: SOP/QA/066 "Procedure for swab/Rinse Sampling". ➤ Cleaning validation protocol of DPI, Ampoule & Liquid vial: 	4	2	2	16					



PHARMA DEVILS

QUALITY ASSURANCE DEPARTMENT

QUALITY RISK ASSESSEMENT AND MITIGATION PLAN

QRA No.:

Name of Facility/Equipment/Utility/System/Activity/Procedure
Risk in reducing Swab & rinse testing

Date of Quality Risk Assessment:

Sr. No.	Item/ Function	Potential Failure Mode (Failure Mode)	Potential Effect of Failure(Effect)	Potential Cause/ Mechanism of Failure	Current Control	Reference	S	O	D	Risk Priority Number (S*O*D)	Recommend- ended Actions (if any)	Post Risk			
												S	O	D	RP N S*O
		Improper sampling	Inattentiveness or Lack of training	<ul style="list-style-type: none"> ➤ Procedure failure. ➤ Product failure. ➤ Contamination. 	<ul style="list-style-type: none"> ➤ Cleaning validation already done for Ampoule ,Liquid Vial & Dry Powder injection. ➤ Cleaning method is robust (Previous trend of swab & rinse). ➤ Training is given as per OJT schedule. 	<ul style="list-style-type: none"> ➤ SOP No.: SOP/QA/066 "Procedure for swab/Rinse sampling". ➤ Cleaning validation protocol of DPI,Ampoule & Liquid vial: ... 									
		Improper Visual inspection	Residue contamination may occur	<ul style="list-style-type: none"> ➤ contamination may pass on to next batch leading to failure in description. ➤ Cleaning agent residue may leads to toxicity ➤ Impurity may pass on to next batch leading to batch failure. 	<ul style="list-style-type: none"> ➤ Cleaning validation already done for Ampoule ,Liquid Vial & Dry Powder injection. ➤ Cleaning method is robust (Previous trend of swab & rinse). ➤ Light intensity of critical areas is verified & recorded during HVAC/Area qualification 	<ul style="list-style-type: none"> ➤ Cleaning validation protocol of DPI,Ampoule & Liquid vial: ... 					Cleaning validation Shall be performed of cleaning agent				



PHARMA DEVILS

QUALITY ASSURANCE DEPARTMENT

QUALITY RISK ASSESSEMENT AND MITIGATION PLAN

QRA No.:

Name of Facility/Equipment/Utility/System/Activity/Procedure

Risk in reducing Swab & rinse testing

Date of Quality Risk Assessment:

Sr. No.	Item/ Function	Potential Failure Mode (Failure Mode)	Potential Effect of Failure(Effect)	Potential Cause/ Mechanism of Failure	Current Control	Reference	S	O	D	Risk Priority Number (S*O*D)	Recommend- ended Actions (if any)	Post Risk			
												S	O	D	RP N S*O
2.	Cleaning Verification	Cleaning Verification not done	Worst case not selected	May leads to contamination	<ul style="list-style-type: none"> ➤ Cleaning validation completed for Ampoule ,Liquid Vial & Dry Powder injection.. ➤ Testing done online at shop floor and instrument is dedicated ➤ pH & Conductivity test shall be done by IPQA . 	<ul style="list-style-type: none"> ➤ SOP No.: SOP/QA/066 "Procedure for swab/Rinse sampling". ➤ Cleaning validation protocol of DPI,Ampoule & Liquid vial: 					Cleaning validation to be evaluated periodically (Yearly). SOP of cleaning validation to be revised.				
		Cleaning Verification not effective	<ul style="list-style-type: none"> ➤ Worst case changed ➤ Deviation in cleaning method ➤ Proper evaluation not done 	May leads to contamination	<ul style="list-style-type: none"> ➤ Evaluation of new API done ➤ SOP of cleaning followed and verified by QA 	<ul style="list-style-type: none"> ➤ "Evaluation of new product for cleaning validation" ➤ SOP's of equipment cleaning 									
3.	Swab/Rinse Sampling	Sampling method not adequate	<ul style="list-style-type: none"> ➤ Untrained person ➤ Sampling method not followed. 	<ul style="list-style-type: none"> ➤ May leads to contamination ➤ False results generated 	<ul style="list-style-type: none"> ➤ SOP of sampling already distributed ➤ Sampling done by trained QA personnel 	<ul style="list-style-type: none"> ➤ SOP no.: "Procedure for Swab/Rinse Sampling". 									
		Untrained person	<ul style="list-style-type: none"> ➤ New person ➤ Negligence 	May leads to contamination	<ul style="list-style-type: none"> ➤ Training given to every new joinee. ➤ OJT training given to IPQA personnel. 	<ul style="list-style-type: none"> ➤ Employee training card 									



PHARMA DEVILS

QUALITY ASSURANCE DEPARTMENT

QUALITY RISK ASSESSEMENT AND MITIGATION PLAN

QRA No.:

Name of Facility/Equipment/Utility/System/Activity/Procedure
Risk in reducing Swab & rinse testing

Date of Quality Risk Assessment:

Sr. No.	Item/ Function	Potential Failure Mode (Failure Mode)	Potential Effect of Failure(Effect)	Potential Cause/ Mechanism of Failure	Current Control	Reference	S	O	D	Risk Priority Number (S*O*D)	Recommend- ended Actions (if any)	Post Risk			
												S	O	D	RP N S*O
		Sampling/locati on procedure wrongly selected	<ul style="list-style-type: none"> ➤ Untrained person ➤ New person ➤ Negligence ➤ Sampling procedure not available. 	May leads to contamination	<ul style="list-style-type: none"> ➤ Cleaning validation completed for Ampoule ,Liquid Vial & Dry Powder injection ➤ SOP of cleaning strictly followed and verified by QA. ➤ OJT training given to IPQA personnel. ➤ SOP of sampling is in place 	<ul style="list-style-type: none"> ➤ SOP No.: "Procedure for swab/Rinse sampling". ➤ Cleaning validation protocol of DPI, Ampoule & Liquid vial: 									
		Sample hold for long period	<ul style="list-style-type: none"> ➤ Hold time not defined ➤ Negligence ➤ UV scanner out of service ➤ Workload 	False results generated	<ul style="list-style-type: none"> ➤ Hold time defined in Analytical Method validation Protocol ➤ Testing done online at shop floor and instrument is dedicated 										
4.	SOP not validated	Cleaning method of equipment not validated	<ul style="list-style-type: none"> ➤ Evaluation of new equipment. 	Cross contamination chance	<ul style="list-style-type: none"> ➤ Procedure for new equipment evaluated after installation is in place. ➤ Change control initiated and shared to QA by Engineering Dept. 	<ul style="list-style-type: none"> ➤ "Evaluation of new equipment for cleaning validation" 									



PHARMA DEVILS

QUALITY ASSURANCE DEPARTMENT

QUALITY RISK ASSESSEMENT AND MITIGATION PLAN

QRA No.:

Name of Facility/Equipment/Utility/System/Activity/Procedure

Risk in reducing Swab & rinse testing

Date of Quality Risk Assessment:

Sr. No.	Item/ Function	Potential Failure Mode (Failure Mode)	Potential Effect of Failure(Effect)	Potential Cause/ Mechanism of Failure	Current Control	Reference	S	O	D	Risk Priority Number (S*O*D)	Recommend- ended Actions (if any)	Post Risk			
												S	O	D	RP N S*O
5.	Visual Inspection	Visual inspector not fit for inspection	<ul style="list-style-type: none"> ➤ Weak eye sight of person. ➤ Untrained person. 	Cross contamination chance	Eye sight tested as per schedule	Medical examination certificate.									
6.	Light Intensity	Light intensity not suitable	Area not qualified	Cross contamination chance	Procedure for check light intensity during area qualification is in place.	Validation master plan									
7.	New Formulation	<ul style="list-style-type: none"> ➤ New API not evaluated 	<ul style="list-style-type: none"> ➤ Worst case might change & SOP's need to be revalidation of API evaluation not performed. ➤ Procedure for evaluation of new API not available. 	Cross contamination chance	<ul style="list-style-type: none"> ➤ Provision for new API evaluation in CVMP ➤ Batch Offer sheet verified daily with current list. 	...									
8.	API sticky in nature	Sticky API not identified	New API not evaluated	<ul style="list-style-type: none"> ➤ Cross contamination chance ➤ Product failure. 											



PHARMA DEVILS

QUALITY ASSURANCE DEPARTMENT

QUALITY RISK ASSESSEMENT AND MITIGATION PLAN

QRA No.:

Name of Facility/Equipment/Utility/System/Activity/Procedure Risk in reducing Swab & rinse testing	Date of Quality Risk Assessment:
--	---

Sr. No.	Item/ Function	Potential Failure Mode (Failure Mode)	Potential Effect of Failure(Effect)	Potential Cause/ Mechanism of Failure	Current Control	Reference	S	O	D	Risk Priority Number (S*O*D)	Recommend- ended Actions (if any)	Post Risk			
												S	O	D	RP N S*O
9.	Method of Analysis	Analysis method not proper	SOP not followed	➤ Effectiveness of method	Procedure of AMV SOP is in place	➤ "Reason of revalidation"									

Where: S=Severity; O=Occurrence Probability; D=Detection; Risk category 1-25 RPN is Low risk, 26-50 RPN is Medium Risk, 51-125 RPN is High Risk.

Name of Facility/Equipment/Utility/System/Activity/Procedure/Unit Operation : Media Fill Activity				Date:			
S.No.	Recommended Action	Responsible Person	Target Date of Completion				
1.	NA	NA	NA				
2.	NA	NA	NA				

CAPA:

If required, mention CAPA No.:

Quality Risk Management Team			Reviewed By Head Production Sign & Date	Approved By Head QA Sign & Date
Name	Department	Sign & Date		



PHARMA DEVILS

QUALITY ASSURANCE DEPARTMENT

QUALITY RISK ASSESSEMENT AND MITIGATION PLAN

QRA No.:

Name of Facility/Equipment/Utility/System/Activity/Procedure
Risk in reducing Swab & rinse testing

Date of Quality Risk Assessment:

QUALITY RISK ASSESSEMENT AND MITIGATION SUMMARY REPORT

Name of Facility/Equipment/Utility/System/Activity/Procedure/Unit Operation : Media Fill Activity

Verification of Action Plan: NA

Remarks (if any): The entire above failure mode and their Severity, Occurrence, Detection rating done and RPN No. is found between 1-25. Hence Risk is detected as low which is acceptable.

Verified By
QA
Sign & Date

Approved By
Head QA
Sign & Date