



# **INSTALLATION QUALIFICATION**

# FOR

# **COMPUTER SYSTEM OF**

# **STABILITY-PC**

System Name	STABILITY-PC
System ID	
Location	QUALITY ASSURANCE
Effective Date	



# PHARMA DEVILS QUALITY ASSURANCE DEPARTMENT

# INSTALLATION QUALIFICATION FOR COMPUTER SYSTEM

### TABLE OF CONTENTS

1.	PR	E APPROVALS	3		
2.	SIGNATURE OF EXECUTOR				
3.	RE	VISION HISTORY	4		
4.	OB	JECTIVE	4		
5.	SC	OPE	5		
6.	SYS	STEM DESCRIPTION	5		
7.	RO	LE AND RESPONSIBILITY	6		
8.	RE	FERENCES	7		
9.	DO	CUMENTATION PROCEDURE	7		
10.	QU	ALIFICATION COMPLETION AND APPROVAL	8		
11.	AC	CEPTANCE CRITERIA	8		
12.	INS	STALLATION VERIFICATION TEST	9		
12	2.1	Identification of System Details	9		
12	2.2	Verification of Master Documents			
12	2.3	Verification of Capacity Requirement	11		
12	2.4	Verification of Hardware Components	12		
12	2.5	Verification of Software Components	14		
12	2.6	Verification of Physical and Logical Security Control	16		
12	2.7	Verification of Test Instruments Calibration and Traceability			
12	2.8	Verification of Power Utility	19		
12	2.9	Verification of Environment Condition			
12	2.10	Verification Of Communication Link Between Server To Computer System	21		
12	2.11	Verification of General System Installation			
12	2.12	Verification of Standard Operating Procedure	23		
13.	INS	STALLATION QUALIFICATION TEST STATUS			
14.	DIS	SCREPANCIES HANDLING DURING COMPUTER SYSTEM QUALIFICAT	ION 24		
15.	DIS	SCREPANCY AND CORRECTIVE ACTION FORM			
16.	AB	BREVIATION			
17.	ΑT	TACHMENT SUMMARY			
18.	INS	STALLATION QUALIFICATION SUMMARY & CONCLUSION			
19.	PO	ST APPROVALS			



### 1. PRE APPROVALS

The signature listed below indicates the preapproval of this installation qualification. This approval is joint responsibility of listed functional areas.

	DOCUMENT DEVELOPMENT	SIGN / DATE
Name Designation	: :	

DOCUMENT REVIEW AND APPROVAL					
Sign / Date	:				
Name	:				
Designation	:				
	Engineering				
Sign / Date	:				
Name	:				
Designation	:				
	IT				
Sign / Date	•				
Name	:				
Designation	:				
	Quality Assurance				
DOCUMENT APPROVAL					
Sign / Date	:				
Name	:				
Designation	:				
	Quality Assurance				



### 2. SIGNATURE OF EXECUTOR

All the executer involved in this document have to sign within prescribed format given below.

#### M/s.....

Name	Designation	Signature	Initial	Date

#### M/s.....

Name	Designation	Signature	Initial	Date

### 3. REVISION HISTORY

Date	Supersedes	Reason for Revision

#### 4. OBJECTIVE

The objective of installation qualification is to collect the sufficient data pertaining to Computer System of STABILITY-PC installed after modification at M/s.....and define the qualification requirements and acceptance criteria for the Computer System STABILITY-PC supporting automation of the system. Successful completion of these qualification requirements will provide assurance that the Computer System of STABILITY-PC for the M/s.....was installed successfully.



### 5. SCOPE

This document is applicable to validation for Hardware and Software system of STABILITY-PC installed after modification at M/s.....This installation qualification shall define the documentation, references and acceptance criteria for validation of Hardware and Software system of STABILITY-PC is installed in accordance with the guidelines laid down by the manufacturer of the system.

#### 6. SYSTEM DESCRIPTION

Computer system of STABILITY-PC defines the controlling of Stability chamber connected to the system. The CS software of stability chamber is a communication software for data management. The operator interface is carried out by CS screen. The CS is used to feed required parameters and set points in the system during operation. The system is connected to data server for printing and data backup. The system is secured by IT through password only within the system.



### 7. ROLE AND RESPONSIBILITY

The validation team comprising of representative from each of the following departments should be responsible for overall compliance with this validation plan.

Department	Responsibilities				
	To collect the necessary data for installation qualification activities.				
	$\succ$ To prepare and execute the installation qualification in coordination with				
Validation Agency	engineering, validation and quality assurance team.				
	Comply with regulatory / Guidelines / Standards / validation plan requirements				
	throughout the validation life cycle.				
	> To submit installation qualification for approval.				
Engineering	> To provide the necessary data for installation qualification activities.				
	> To review the installation qualification.				
IT	> To provide the necessary data for installation qualification activities.				
	> To review the installation qualification.				
Quality Assurance	To provide the necessary data for installation qualification activities.				
	$\succ$ To review the installation qualification.				
Quality Assurance	> To approve and authorized the installation qualification.				



### 8. REFERENCES

The publication listed below form part of this reference documents. Each publication shall have latest revision in effect on the date of this document is approved for execution.

GAMP 5	Good Automated Manufacturing Practices, Version 5, GuidelineDocument for Automated Systems from International Society ofPharmaceutical Engineering
21 Code of Federal Regulations (CFR), Part 210	Current Good Manufacturing Practice in Manufacturing, Processing, Packing, or Holding off Drugs; General
21 Code of Federal Regulations (CFR), Part 211	Current Good Manufacturing Practice for finished Pharmaceuticals
21 Code of Federal Regulations (CFR), Part 11	<ul><li>21 Code of Federal Regulations (CFR), Part 11</li><li>Electronic Records, Electronic Signatures, Final Rule Electronic</li><li>Submissions; Establishment of Public Docket, Notice</li></ul>
ICH Q9	International Conference of Harmonization (ICH) quality risk assessment Q9
EU GMP	Laying down the principles and guidelines of GMP in respect of medicinal products for human use.
SRS	System Requirement Specification
WHO	Appendix 5, validation of computerized systems.

#### 9. DOCUMENTATION PROCEDURE

- Qualification activities will be performed as defined in the approved document.
- All documentation will be completed during the execution of the qualification.
- Recording of information will be made in permanent ink.
- Fill out complete information in the verification table provided.
- Do not keep any space blank. Mark blank space with a single line throughout the appropriate space with mentioning NA (Not Applicable) and put initial and date.
- Correct the mistakes by drawing a single line through the incorrect data, recording the correct information and then initialing and dating the change.



### 10. QUALIFICATION COMPLETION AND APPROVAL

- Verify that all tests required by qualification are completed and attached.
- Verify that all amendments and discrepancies are documented, approved and attached.
- If all items in the qualification for the validation to Hardware and Software system of STABILITY-PC been reviewed and found to be acceptable, sign the corresponding block in the qualification completion and approval form.

#### **11. ACCEPTANCE CRITERIA**

- Installation completion as per manufacturer's recommendations & cGMP requirements.
- Installation of major components as per the design specifications.
- The supply of all necessary documentation from manufacturer.



### 12. INSTALLATION VERIFICATION TEST

### **12.1** Identification of System Details

Objective	:	This test sheet is intended to verification of equipment details.
Tools Required	:	Not Applicable
Procedure	:	1. Record Equipment Name
		2. Record Identification No.
		3. Record Equipment Location
Acceptance	:	Data recorded from the equipment shall match with the data specified in
Criteria		verification table.

### Verification Table:

Equipment Details	Specified As	As observed	Discrepancy? (Y/N)
Equipment Name	STABILITY-PC		
Identification No.			
Location	QUALITY ASSURANCE		

] No

Remarks:

Meet the acceptance Criteria	[	] Yes	[
------------------------------	---	-------	---

Checked by :\_\_\_\_\_

Verified by : \_\_\_\_\_

Date: \_\_\_\_\_

Date:



#### **12.2** Verification of Master Documents

Objective	:	To verify the availability of related master documents.
Tools Required	:	Not Applicable
Procedure	:	1. Verify Documents Name.
		2. Verify Documents Reference.
		3. Verify Documents Availability.
Acceptance	:	Documents should be available.
Criteria		

#### Verification Table:

Documents Name	Documents Reference	Availability (Yes/No)	Verified (Yes/No)	Discrepancy? (Y/N)
SRS	-			
Operational Manual	-			
BOM	Refer attachment No. 1			

Remarks:

Meet the acceptance Criteria [ ] Yes [ ] No

Checked by :\_\_\_\_\_

Verified by :\_\_\_\_\_

Date: \_\_\_\_\_

Date:\_\_\_\_\_



### 12.3 Verification of Capacity Requirement

Objective	: To verify the processing capacity of Computer System
Tools required	: Not Applicable
Procedure	: Physical verification of Capacity Requirement as per SRS
	1. No more than 50% of the installed hard disk capacity in PC
	components should be consumed by installed software.
	2. Historical data storage capacity should allow for online retrieval
	of at forever of any historical data.
1	: 1. Capacity Requirement of the control system shall match with
criteria	SRS.

#### Verification Table:

S. No.	Item Name	Expected	Actual (Yes/No)	Discrepancy? (Y/N)
1	Processing Power	Not more than 50% of the processing capacity		
2	Memory	Not more than 50% of the installed physical memory		
3	Local Electronic Storage	Not more than 50 % should be consumed by installed software		

#### Remarks:

 Meet the acceptance Criteria [
 ] Yes [
 ] No

 Checked by :
 \_\_\_\_\_\_\_
 Date:
 \_\_\_\_\_\_\_

 Verified by :
 \_\_\_\_\_\_\_\_
 Date:
 \_\_\_\_\_\_\_\_



### 12.4 Verification of Hardware Components

Objective	:	To verify the installed hardware components as per the SRS.
Tools Required	:	Not Applicable
Procedure	:	1. Verify Hardware Name.
		2. Verify Hardware Make/ Assemble By
		3. Verify Hardware Model No./Specification
Acceptance	:	Installed hardware component should match with SRS.
Criteria		

#### **Verification Table:**

Name	Make/ Assemble By	Model No./Specification	Qty.	Installation (Yes/No)	Discrepancy? (Y/N)
Monitor	Acer	V196HQC	01		
CPU	Acer	Veriton-IE4066	01		
Keyboard	Acer	NA	01		
Mouse	Acer	NA	01		
UPS	Emerson Network System	20 kVA	01		
RAM	Acer	4 GB	01		
Processor	Intel	Pentium(R) G3260 3.30 GHz	01		
Printer	Canon	3300	01		
Stability Chamber	Thermolab scientific equipmants	SG1K00A501	03		

### Verification Table: Communication Port

S. No.	Port Type	Qty.	Installation (Yes/No)	Discrepancy? (Y/N)
1	USB	2		
2	Ethernet	2		

Remarks:

Meet the acceptance Criteria [ ] Yes [ ] No





Checked by :\_\_\_\_\_

Date:

Verified by :\_\_\_\_\_

Date:\_\_\_\_\_



### 12.5 Verification of Software Components

Objective	:	To verify the installed software components as per the SRS.
Tools Required	:	Not Applicable
Procedure	:	1. Verify Software Name.
		2. Verify Software Version
		3. Verify operating system.
		4. Verify software backup availability.
Acceptance Criteria	:	Installed software component should match with SRS.

#### Verification Table:

#### A. For Software Components

S.No.	Software Name	Version	Installation (Yes/No)	Discrepancy? (Y/N)
1	Adobe Reader XI	11.0.07		
2	Microsoft Office Enterprises 2007	12.0.4518.1014		
3	Mozilla Firefox	62.0		
4	StabilitySP2For1.2	1.0.0		
5	WinRAR 5.50	5.50.0		
6	Windows 7 Professional	7 2009		

### **B.** Operating system details

S.No.	Window	Product key/ Liecence key	Discrepancy? (Y/N)
1	Windows 7		

### C. Software Backup Availability

S.No.	Available (Yes/ No)	Discrepancy? (Y/N)





Remarks:

Meet the acceptance Criteria [ ] Yes [ ] No	
Checked by :	Date:
Verified by :	Date:



### 12.6 Verification of Physical and Logical Security Control

Objective	:	Verify the physical and logical security of Computer System.
Tools Required	:	Not Applicable
Procedure	:	1. Verify physical Security.
		2. Verify logical security of Application Window and Software
		3. Verify User for Application access
Acceptance	:	Physical security should be maintained. Logical
Criteria		Window Login password should be available.
		Application Software should have multiple numbers of user's role
		with user name.

### A. Verification Table for Physical Security:

System	Security	Availablility (Yes/No)	Discrepancy? (Y/ N)
Computer system	PCB should be secure in control Panel.		

### **B.** Verification Table for Logical Security Window:

Specified user	Logical security available (Yes/No)	Discrepancy? (Y/ N)
Admin		
User		

#### C. Verification Table for Logical User Application Software:

Specified user	Logical User available (Yes/No)	Discrepancy? (Y/ N)
Admin		
Master		
User		

Remarks:



# PHARMA DEVILS QUALITY ASSURANCE DEPARTMENT

# **INSTALLATION QUALIFICATION FOR COMPUTER SYSTEM**

Meet the acceptance Criteria [	] Yes [	] No	
Checked by :			Date:
Verified by :			Date:



### 12.7 Verification of Test Instruments Calibration and Traceability

Objective	:	To verify the test instruments traceability.
Tools Required	:	Not Applicable
Procedure	:	1. Verify Certificate No.
		2. Verify Traceability.
		3. Verify Calibration Done Date.
		4. Verify Calibration Due Date.
Acceptance	:	Test instruments should be calibrated at the execution.
Criteria		

#### Verification Table:

Certificate No.	Traceability	Calibration Done On	Calibration Due On	Verified (Yes/No)	Discrepancy? (Y/N)
	Refer attachment No. 2				
	Refer attachment No. 3				

Remarks:

Meet the acceptance Criteria [ ] Yes [ ] No

Checked by : \_\_\_\_\_

Verified by : \_\_\_\_\_

Date:\_\_\_\_\_

Date:\_\_\_\_\_



### **12.8** Verification of Power Utility

Objective	:	To verify the installed power utility is as per specifications.
Tools Required	:	Digital Multimeter
Procedure	:	1. Switch ON the Power Supply.
		2. Put the Multimeter in AC/DC range.
		3. Record the supply voltage.
Acceptance	:	Measured voltage shall match with the specified voltage.
Criteria		

#### Verification Table:

#### **Supply Voltage Measurement:**

Name	Specified Voltage	Measured Voltage	Discrepancy? (Y/N)
Computer System Power Supply	230 VAC		
Stability Chamber	(220-240) VAC		

Remarks:

Meet the acco	eptance Criteria [	] Yes [	] No		
Checked by	:			Date:	
Verified by	:			Date:	



#### **12.9** Verification of Environment Condition

Objective	:	To verify the environment conditions.
Tools Required	:	Digital Thermo Hygrometer
Procedure	:	1. Switch ON the thermo hygrometer.
		2. Record maximum temperature.
		3. Record maximum relative humidity.
Acceptance	:	Test instruments should be calibrated at the execution.
Criteria		

### Verification Table:

Name	Temperature	Relative Humidity	Measured Results	Discrepancy? (Y/N)
Computer System Environmental Condition	NMT 25 °C	NA		
Stability chamber	NMT 25 °C	NA		

#### Remarks:

	Meet the acceptance	Criteria	[	] Yes	[	] No
--	---------------------	----------	---	-------	---	------

Checked by :\_\_\_\_\_

Verified by :\_\_\_\_\_

Date:\_\_\_\_\_

Date:\_\_\_\_\_



### 12.10 Verification Of Communication Link Between Server To Computer System

Objective	:	To Verify the communication link with server to client
Tools Required	:	Not Applicable
Procedure	:	1. Verify and record the communication link between PC to Server
		2. Verify and record the communication link between PC to Printer.
Acceptance	:	Communication link ping with PC to server should be executed and report
Criteria		should be proper.

### Verification Table:

S.No.	Source	Destination	Ping Executed (Yes/No)	Discrepancy? (Y/N)
01	Computer system	File Server		
02	Computer system	Printer		

Remarks:

Meet the acceptance Criteria [	] Yes [	] No	Refer Attachment No. [ ]
Checked by :			Date:
Verified by :			Date:



### 12.11 Verification of General System Installation

Objective	:	To verify the general system installation.
Tools Required	:	Not Applicable
Procedure	:	1. Check all the test given in verification table.
		2. Record the result in verification table.
Acceptance	:	All the test result should match with expected result.
Criteria		

#### Verification Table:

Description	Expected Result (Yes/No)	Discrepancy? (Y/N)
Major components should be protected from shock.		
No visible physical damage should be available.		
Sufficient space should be available for maintenance.		
System identification nameplate should be available.		
System should be installed with all necessary instruments.		
Earthling should be connected properly.		
Power and signal cable should be separate.		
Unterminated and broken wire should not be open.		

Remarks:

Meet the acceptance Crit	eria [	] Yes	[	] No
where the deceptunce of a		] 100	L	1110

Checked by	:
------------	---

Verified by :\_\_\_\_\_

Date:\_\_\_\_\_

Date:



## 12.12 Verification of Standard Operating Procedure

Objective	:	To verify the availability of related standard operating procedure.
Tools Required	:	Not Applicable
Procedure	:	1. Verify SOP Name.
		2. Verify SOP No.
		3. Verify SOP Availability.
Acceptance Criteria	:	Documents should be available.

### Verification Table:

SOP Name	SOP No.	Availability (Yes/No)	Discrepancy ? (Y/N)
Stability study			
Remarks: Meet the acceptance Criteria [ Checked by : Verified by :		Date :	



### 13. INSTALLATION QUALIFICATION TEST STATUS

The installation qualification test status is as per below mentioned table.

Test Description	Status (Pass / Fail)	Discrepancy? (Y/N)
Identification of System Details		
Verification of Master Documents		
Verification of Capacity Requirement		
Verification of Hardware Components		
Verification of Software Components		
Verification of Physical and Logical Security Control		
Verification of Test Instruments Calibration and Traceability		
Verification of Power Utility		
Verification of Environmental Condition		
Verification Of Communication Link Between Server To Computer		
System		
Verification of General System Installation		
Verification of Standard Operating Procedures		

### 14. DISCREPANCIES HANDLING DURING COMPUTER SYSTEM QUALIFICATION

- In case of discrepancy observed during qualification, document in the defined column in each table and document the details of the observation in the discrepancy log sheet.
- Inform to User, engineering and quality assurance about discrepancy.
- Investigate the discrepancy and ensure the possible impact.
- If discrepancy does not have potential to impact on operation as well as performance of the system, close the discrepancy with proper justification.
- The User, engineering and QA will decide whether discrepancy is acceptable or not.
- If discrepancy is acceptable, provide conclusion and recommendation if any into respective column.



PHARMA DEVILS QUALITY ASSURANCE DEPARTMENT

# **INSTALLATION QUALIFICATION FOR COMPUTER SYSTEM**

### **15. DISCREPANCY AND CORRECTIVE ACTION FORM**

Protocol Reference	
Discrepancy Number	

#### DISCREPANCY

Describe the Discrepancy		
Reported by	Date	

#### **CORRECTIVE ACTION**

Describe corrective action taken (Attach additional sheets if necessary)		
Reported by	Date	

#### **DISPOSITION ACTION**

Acceptable?	Yes	No		
Discussion				
Approved by			Date	
COMPLETION				

Completed by	Date

#### **16. ABBREVIATION**

Abbreviations	Description



# PHARMA DEVILS QUALITY ASSURANCE DEPARTMENT

# **INSTALLATION QUALIFICATION FOR COMPUTER SYSTEM**

GMP	Good Manufacturing Practices	
IQ	Installation Qualification	
OQ	Operation Qualification	
QA	Quality Assurance	
SOP	Standard Operating Procedure	
NA	Not Applicable	
ICH	International Conference of Harmonization	
mA	Mili Ampere	
VAC	Alternate Current Voltage	
VDC	Direct Current Voltage	
RH	Relative Humidity	
CS	Computer System	
NMT	Not More Than	





### **17. ATTACHMENT SUMMARY**

Attachment No.	Description

### **18. INSTALLATION QUALIFICATION SUMMARY & CONCLUSION**

Compiled by: \_\_\_\_\_

Date:



### **19. POST APPROVALS**

The signature listed below indicates the post approval of this installation qualification. This approval is joint responsibility of listed functional areas.

	DOCUMENT DEVELOPMENT	SIGN / DATE
Name	:	
Designation	:	

DOCUMENT REVIEW AND APPROVAL			
Sign / Date	:		
Name	:		
Designation	:		
	Engineering		
Sign / Date	:		
Name	:		
Designation	:		
	IT		
Sign / Date	:		
Name	:		
Designation	:		
	Quality Assurance		
DOCUMENT APPROVAL			
Sign / Date	:		
Name	:		
Designation	:		
	Quality Assurance		