



STANDARD OPERATING PROCEDURE

Department: Microbiology	SOP No.:
Title: Cleaning of Glasswares for Bacterial Endotoxin Test	Effective Date:
Supersedes: Nil	Review Date:
Issue Date:	Page No.:

1.0 OBJECTIVE:

To lay down a procedure for Cleaning of Glasswares for Bacterial Endotoxin Test.

2.0 SCOPE:

This SOP is applicable for Cleaning of Glasswares for Bacterial Endotoxin Test in Microbiology Lab of Quality Control Area.

3.0 RESPONSIBILITY:

Officer / Executive – Microbiology

4.0 ACCOUNTABILITY:

Head – QC

5.0 ABBREVIATIONS:

°C	Degree Celsius
BET	Bacterial Endotoxin Test
Hrs.	Hours
Ltd.	Limited
ml	Milliliter
No.	Number
QA	Quality Assurance
QC	Quality Control
SOP	Standard Operating Procedure
WFI	Water for Injection

6.0 PROCEDURE:

- 6.1 Dip the Test Tubes to be cleaned in 1000 ml of Beaker & add sufficient amount of purified water.
- 6.2 Remove the gel content from used test tubes with brush and purified water.
- 6.3 After cleaning of test tubes with brush rinse with purified water.
- 6.4 New non Depyrogenated test tubes rinse with purified water.
- 6.5 After cleaning, keep the tubes in oven for drying.

7.0 ANNEXURES:

NA

ENCLOSURES: SOP Training Record.

8.0 DISTRIBUTION:



PHARMA DEVILS
MICROBIOLOGY DEPARTMENT

STANDARD OPERATING PROCEDURE

Department: Microbiology

SOP No.:

Title: Cleaning of Glasswares for Bacterial Endotoxin Test

Effective Date:

Supersedes: Nil

Review Date:

Issue Date:

Page No.:

- Controlled Copy No. 01 Quality Assurance
- Controlled Copy No. 02 Microbiology
- Master Copy Quality Assurance

9.0 REFERENCES:
In House

10.0 REVISION HISTORY:

CHANGE HISTORY LOG

Revision No.	Change Control No.	Details of Changes	Reason for Change	Effective Date	Updated By