



PHARMA DEVILS

ENGINEERING DEPARTMENT

STANDARD OPERATING PROCEDURE

Department: Engineering	SOP No.:
Title: Specification of RTD Sensors	Effective Date:
Supersedes: Nil	Review Date:
Issue Date:	Page No.:

1.0 Revision History

Rev. No.	Details of changes	Reason for change
00	NIL	NEW SOP

2.0 OBJECTIVE:

2.1 The Objective of this SOP is to provide the Specifications of RTD sensor.

3.0 SCOPE:

3.1 This SOP is the provide the Specifications of RTD sensor.

4.0 RESPONSIBILITY:

4.1 The Maintenance Engineer shall be:

4.1.1 Responsible for ensuring the Specifications of RTD Sensors at the time of releasing new order.

4.1.2 Responsible to ensure the compliance of Specifications of RTD Sensors at the time of receipt.

5.0 ACCOUNTABILITY:

Head –Engineering Services

6.0 PROCEDURE:

6.1 RTD Sensors:

6.1.1 Class : PT-100

6.1.2 Type : Three wire

6.1.3 Temperature range : 0-400 Degree Centigrade.

6.1.4 Application : Temperature monitoring / Validation.



PHARMA DEVILS

ENGINEERING DEPARTMENT

STANDARD OPERATING PROCEDURE

Department: Engineering	SOP No.:
Title: Specification of RTD Sensors	Effective Date:
Supersedes: Nil	Review Date:
Issue Date:	Page No.:

7.0 ANNEXURES:

NIL

8.0 References (S)

NIL

9.0 Glossary

SOP : Standard Operating procedure
No : Number
RTD : Resistance Temperature Detector