



PHARMA DEVILS

QUALITY ASSURANCE DEPARTMENT

FAILURE MODE EFFECT ANALYSIS FOR PROCESSING OF EYE DROP

S.No.	Item/ Function	Potential Failure Mode (Failure Mode)	Potential Effect of Failure (Effect)	Potential Cause/ Mechanism of Failure	Current Control	Reference	S	O	D	Risk Priority Number (S*O*D)	Recommend- ended Actions (if any)	Post Risk			
												S	O	D	RP N S*O
Product Processing:															
1.	Qualified Packers	<ul style="list-style-type: none"> ➤ Packers are not qualified to check the overprinting of the labels. 	<ul style="list-style-type: none"> ➤ In case of wrong or no overprinting on vial / ampoule there might be a chance of packing of without printed labels, which may cause may leads to market complaint in future. ➤ Incorrect or illegible printed matter will mislead the patient lead to market complaint. 	<ul style="list-style-type: none"> ➤ Non-qualified visual inspectors verify the vial/ampoule coding. ➤ Casual approach of visual inspectors during visual inspection. 	<ul style="list-style-type: none"> ➤ During labeling operation of vial/ampoules, visual inspectors continuously verify the vial/ampoules coding. ➤ The visual inspectors are rotate after every two hours. ➤ SOP of sticker labeling is available. 	As per SOP.	3	2	1	6 Low category & Risk Accepted	Adequate procedure no recommendation required.	N A	NA	NA	N A
2.	Label sensors	<ul style="list-style-type: none"> ➤ Presence of label can't be checked automatically during labeling. 	<ul style="list-style-type: none"> ➤ Unlabelled and mislabeled vials /ampoules will mix with good ampoule/vial. ➤ Product may be produced with incomplete information. ➤ There might be a chance of packing of without labeled vial/ampoule followed by market complaint in future. 	<ul style="list-style-type: none"> ➤ "No label" sensor is available to check the presence or absence of label. 	<ul style="list-style-type: none"> ➤ "No label" sensor is available to check the presence or absence of the label during labeling. ➤ At starting of batch or any after any break the sensor is challenged and recorded the details in its log book. 	As per SOP.	3	2	1	6 Low category & Risk Accepted	Adequate procedure no recommendation required.	N A	NA	NA	N A



PHARMA DEVILS

QUALITY ASSURANCE DEPARTMENT

FAILURE MODE EFFECT ANALYSIS FOR PROCESSING OF EYE DROP

S.No.	Item/ Function	Potential Failure Mode (Failure Mode)	Potential Effect of Failure (Effect)	Potential Cause/ Mechanism of Failure	Current Control	Reference	S	O	D	Risk Priority Number (S*O*D)	Recommend- ended Actions (if any)	Post Risk			
												S	O	D	RP N S*O
3.	Air Filters	<ul style="list-style-type: none"> ➤ Filter Chocking / leakage ➤ Filter integrity test not possible. 	<ul style="list-style-type: none"> ➤ Without filtered air directly impacted to Product Sterility & Quality. ➤ Product contaminated. ➤ Environmental contamination of product 	<ul style="list-style-type: none"> ➤ Filter use more than recommended cycle ➤ Filter Use without Integrity testing. 	<ul style="list-style-type: none"> ➤ Filter use as per recommended cycle & maintained in log book for their cycle. ➤ Air filters filter integrity done according to the frequency; Manufacturing tank- Monthly, Holding Tank: - 15 days, Buffer Vessel: - 15 Days, Filling Machine: - Daily, Pendants: - Monthly. ➤ Pressure gauge available to monitor differential pressure across filter. 	As per SOP.	4	2	1	8 Low category & Risk Accepted	Adequate procedure no recommendation required.	N A	NA	NA	N A
4.	Nitrogen Gas	<ul style="list-style-type: none"> ➤ Nitrogen gas quality failure. 	<ul style="list-style-type: none"> ➤ Particle and microbial contamination of Vials may leads to product failure. 	<ul style="list-style-type: none"> ➤ No provision to filter the nitrogen gas before use. ➤ Sampling and testing of nitrogen for sterility testing is not done periodically. ➤ Purity of nitrogen is not identified. ➤ Nitrogen gas generator didn't qualified. 	<ul style="list-style-type: none"> ➤ Nitrogen gas supply to the equipment has been qualified & COA available for purity. ➤ Sampling and testing for sterility of nitrogen gas is done periodically, As PER SOP. ➤ Before pre-purging and post purging station two dedicated 0.2 μ filters are in place and one 0.2 μ filter is installed at pendent in filling room. ➤ SOP of Integrity checking of Nitrogen filter in place. 	As per SOP.	4	1	1	4 Low category & Risk Accepted	Adequate procedure no recommendation required. Hence risk is accepted.	N A	N A	N A	N A



PHARMA DEVILS

QUALITY ASSURANCE DEPARTMENT

FAILURE MODE EFFECT ANALYSIS FOR PROCESSING OF EYE DROP

S.No.	Item/ Function	Potential Failure Mode (Failure Mode)	Potential Effect of Failure (Effect)	Potential Cause/ Mechanism of Failure	Current Control	Reference	S	O	D	Risk Priority Number (S*O*D)	Recommend- ended Actions (if any)	Post Risk			
												S	O	D	RP N S*O
5.	Power failure	<ul style="list-style-type: none"> ➤ Critical processes stops, like LAF over filling machine, blower of tunnel, HVAC system etc. 	<ul style="list-style-type: none"> ➤ Area gets contaminated due to stopped HVAC system. ➤ Vials cannot be depyrogenated properly due to shut down of blower. ➤ Filling area gets contaminated due to LAF stopped. ➤ Process out of specification. ➤ Unsafe if start automatically on restoration of power 	<ul style="list-style-type: none"> ➤ Power shut down. ➤ No electricity. ➤ Technical fault. ➤ No UPS supply available. ➤ No DG set available. 	<ul style="list-style-type: none"> ➤ Machine is not start automatically without operator intervention after incident. ➤ Provision of UPS is also there to the control the system in case of power failure. ➤ LAF Over the filling machine, blower of tunnel, HMI of all machines have provision of UPS. ➤ 02 DG sets are available which can be started within 30 sec. ➤ Area Recovery study has been performed. ➤ SOP FOR Procedure to be followed in case of power failure is in place. 	As per SOP.	3	2	2	12 Low category & Risk Accepted	Adequate procedure no recommendation required. Hence risk is accepted.	N A	N A	N A	N A

Where: S=Severity; O=Occurrence Probability; D=Detection; Risk category 1-25 RPN is Low risk, 26-50 RPN is Medium Risk, 51-125 RPN is High Risk.

Name of Facility/Equipment/Utility/System/Activity/Procedure/Unit Operation: Product Processing of Eye Drop			Date:		
S. No.	Recommended Action	Responsible Person	Target Date of Completion		
1.	NA	NA	NA		
2.	NA	NA	NA		

CAPA: Not required



PHARMA DEVILS
QUALITY ASSURANCE DEPARTMENT

FAILURE MODE EFFECT ANALYSIS FOR PROCESSING OF EYE DROP

If required, mention CAPA No.: NA

Quality Risk Management Team			Reviewed By Head Operations Sign & Date	Approved By Head QA Sign & Date
Name	Department	Sign & Date		

QUALITY RISK ASSESSEMENT AND MITIGATION SUMMARY REPORT

Name of Equipment: Product Processing of Eye Drop

Verification of Action Plan: NA

Remarks (if any): The entire above failure mode and their Severity, Occurrence, Detection rating done and RPN No. is found between 3 to 12. Hence Risk is detected as low which is acceptable.

Verified By
QA
Sign & Date

Approved By
Head QA
Sign & Date